


Test of "SPLASH RESISTANCE" in agreement with the guidelines of ISO 22609:2004(E). Client PFC Pharma Manufacturing SL, mask reference code ID2

**Report n°: TOP2020_R81
Edition: 01
Page: 1 di 6**

CLIENT	PFC Pharma Manufacturing SL Calle Osona, 2 08820, Prat de Llobregat (El), Barcelona Cataluña, España		
LABORATORY	<input type="checkbox"/> MaB – Microscopia applicata e biologia cellulare - <input checked="" type="checkbox"/> ToP – Tossicologia e Proteomica - <input checked="" type="checkbox"/> Ms ² – Materiali, sensori e sistemi -		
Analysis conducted by: Alberto Ferrari Alberto.Ferrari@tpm.bio	Signature 	Date	26/05/2020
Head of the laboratory: Aurora Cuoghi, PhD aurora.cuoghi@tpm.bio	Signature 	Date	27/05/2020
Approved by: Aldo Tomasi, aldo.tomasi@unimore.it Scientific Director of Toxicology and Proteomic Laboratory	Signature 	Date	27/05/2020

Ed.	Report n°	Date	Description
01	TOP2020_R81	27/05/2020	First edition

Test of "SPLASH RESISTANCE" in agreement with the guidelines of ISO 22609:2004(E). Client PFC Pharma Manufacturing SL, mask reference code ID2 Report n°: **TOP2020_R81**
Edition: 01
Page: 2 di 6

Index

1. ORDER REFERENCE	3
2. PURPOSE.....	3
2.1 SPECIMEN.....	3
2.2 Sample preparation.....	3
3. MATERIALS & METHODS.....	3
3.1 MATERIALS	3
3.2 Instrumentations	4
3.3 Experimental method	4
3.4 Experimental condition:	4
3.5 Acceptance criterion	5
4. RESULTS	5
5. CONCLUSIONS	6

Test of "SPLASH RESISTANCE" in agreement with the guidelines of ISO 22609:2004(E). Client PFC Pharma Manufacturing SL, mask reference code ID2

**Report n°: TOP2020_R81
Edition: 01
Page: 3 di 6**

1. ORDER REFERENCE

TPM_2020_780_BIO

2. PURPOSE

"SPLASH RESISTANCE" analysis: evaluation of the resistance of the device to the penetration of a certain volume of synthetic blood by high-speed impact between liquid and device for a short period of time (1 second). The analysis is carried out following the guidelines of ISO 22609:2004(E).

2.1 SPECIMEN

Perfect Care Distribution srl supplied to the laboratory 32 complete face masks, from the production batch FM2003. Mask code **ID2**, denomination 3ply-adult.

2.2 Sample preparation

Samples have been tested without any modification in their geometry, whatsoever. The sample is pre-conditioned in a climatic chamber at a temperature of 21 ° C and relative humidity of 85% for 4 hours before the analysis. The measurement is made within 1 minute of removal from the climatic chamber.

3. MATERIALS & METHODS

3.1 MATERIALS

- Demineralized H₂O 0.055 µS / cm
- Triton X 100 X Sigma-Aldrich cod. T8787; batch MKBR5267V
- Direct RED 80 sigma aldrich cod. 365548; batch MKBB6842V

Synthetic blood is made from a 15 mg / L solution of Triton X 100 and a Direct RED 80 red color 200 mg / L in demineralized water.

Fondazione Democenter-Sipe

via P. Vivarelli 2, 41125 – Modena • P.I. e C.F. 01989190366 • Tel. +39 059 2058146 - Fax +39 059 2058161
info@fondazionedemocenter.it • democentersipe@pcert.it • www.democentersipe.it

Tecnopolo di Mirandola

via 29 Maggio 6, 41037 Mirandola (Mo) • www.tpm.bio • Tel +39 0535 613801

FormTPM 010E Rev04 del 12.04.2018

Test of "SPLASH RESISTANCE" in agreement with the guidelines of ISO 22609:2004(E). Client PFC Pharma Manufacturing SL, mask reference code ID2

**Report n°: TOP2020_R81
Edition: 01
Page: 4 di 6**

3.2 Instrumentations

- "Flower 340" climatic chamber Serial Number: 011TT29. Performance certificate valid until September 2020.
- "Winkratos 5.00" software.
- 3D-Bioplotter ENVISIONTEC, serial number ETB41507M056, (MS2_066)

3.3 Experimental method

The analysis is based on the visual observation of the sample subjected to a squirt of synthetic blood at high speed to simulate an accidental leakage of the patient's blood in the surgical site. The sample is mounted on a special support perpendicular to the direction of the liquid flow. The squirt of synthetic blood, whose speed and quantity are comparable to the excision of a large artery, takes place by pneumatic impulse through a syringe containing synthetic blood, a needle of defined section and length and a piston on which electronically regulated pressure is exerted via software. The quantity of liquid dispensed is 2.0 ml. The observation is done visually and through the use of a tissue paper, noting that the liquid does not pass through the mask or does not wet the inside after 10 seconds from performing the test. Synthetic blood is prepared using a solution of Triton X 100 in order to have a surface tension of 0.042 N / m, comparable to that of whole blood.

3.4 Experimental condition:

The experimental parameters for the test have been set as indicated below:

Sample-cannula distance	Cannula internal diameter	Cannula length	Pressure	Pulse duration
30 cm	0.84 mm	12.7 mm	21 kPa	0.7 s

Test of "SPLASH RESISTANCE" in agreement with the guidelines of ISO 22609:2004(E). Client PFC Pharma Manufacturing SL, mask reference code ID2 Report n°: **TOP2020_R81**
Edition: 01
Page: 5 di 6

3.5 Acceptance criterion

The test is performed on 32 different masks. To have an AQL of 4% the test is considered passed if the number of samples that show the resistance to penetration of liquid are at least 29.

4. Results

The masks with sample code **ID2** have been subjected to pre-treatment and splash resistance test. Figure 1 shows a representative image of the internal and external part of a sample template.




Figure 1: External side at the top and the internal side at the bottom, after splash resistance test

Fondazione Democenter-Sipe

via P. Vivarelli 2, 41125 – Modena • P.I. e C.F. 01989190366 • Tel. +39 059 2058146 - Fax +39 059 2058161
info@fondazionedemocenter.it • democentersipe@pcert.it • www.democentersipe.it

Tecnopolo di Mirandola

via 29 Maggio 6, 41037 Mirandola (Mo) • www.tpm.bio • Tel +39 0535 613801

FormTPM 010E Rev04 del 12.04.2018

Test of "SPLASH RESISTANCE" in agreement with the guidelines of ISO 22609:2004(E). Client PFC Pharma Manufacturing SL, mask reference code ID2 **Report n°: TOP2020_R81**
Edition: 01
Page: 6 di 6

None of the 32 masks shows the permeation of synthetic blood in the internal part of the mask within 10 seconds, and also in greater times, from the application of the squirt of liquid at the pressure of 21kPa. All 32 replicates show the same performance.

5. Conclusions

The tests carried out indicate that the materials used may be suitable for the construction of a mask classifiable as IIR.